

The New International Temperature Scale of 1990 and its Effect on Radiometric, Photometric, and Colorimetric Measurements and Standards

K. D. Mialenz, R. D. Saunders, A.C. Parr and J. J. Hiss
National Institute of Standards and Technology
Gaithersburg, MD 20899
U.S.A.

ABSTRACT

The differences between ITS-90 and IPTS-68 above 1235K are described. It is shown that none of the following CIE definitions or recommendations require revision because of the introduction of the ITS-90: International Lighting Vocabulary definitions; CIE Standard Illuminants A, D₆₅, other illuminants; and sources for realizing CIE illuminants. The effect of the ITS-90 on previously calibrated sources for realizing CIE illuminants is negligibly small.

Résumé

Les différences entre ITS-90 et IPTS-68 au delà de 1235K sont décrites. Il est montré qu'aucune des suivantes définitions ou recommandations de la CIE nécessite de révision à cause de l'introduction du ITS-90: Les définitions du Vocabulaire Internationale de l'Eclairage, les Illuminants Standards A, D₆₅, et autres illuminants; et les sources pour la réalisation d'illuminants CIE. L'effet du ITS-90 sur les sources précédemment calibrées pur réaliser les illuminants CIE est négligeablement faible.

Zusammenfassung

Nach einer Diskussion der Unterschiede zwischen der ITS-90 und der IPTS-68 im Temperaturbereich ueber 1235 K wird gezeigt, dass die Einfuehrung der ITS-90 keine Revision der folgenden CIE Definitionen oder Empfehlungen erforderlich macht: Definitionen des Internationalen Wörterbuchs der Lichttechnik; Normlichtarten A, D₆₅, und andere; und CIE Normlichtquellen. Der Einfluss der ITS-90 auf bereits geeichte Normlichtquellen ist vernachlässigbar klein.

1. The 1990 Gold-Point Change

The radiometric and photometric measurement scales of many national laboratories are based on blackbody physics. For example, the NIST calibration services in thermal radiometry are based on measurement scales derived from blackbody physics. A blackbody radiator at the temperature of the freezing point of gold and Planck's radiation law are used to establish the NIST radiance-temperature and spectral-radiance scales by calibrating a variable-temperature blackbody against the gold-point blackbody at 654.6 nm and performing subsequent measurements of

spectral-radiance ratios to extrapolate this calibration to extended temperature and spectral ranges. The spectral-radiance scale is then used to derive the NIST scale of spectral irradiance by a radiance-to-irradiance transfer, and hence the NIST scale of luminous intensity, luminous flux, and color temperature are derived by spectral-irradiance calibrations of photometer lamps and computations of these quantities according to the standard procedures established by the Commission International de l'Eclairage (CIE). All of these steps have been documented in NIST publications [1-4]. From 1968 until June 30, 1990, the temperature of the primary blackbody standard use in these scale realizations was that assigned to the freezing point of gold in the International Practical Temperature scale of 1968 (IPTS-68) [5].

$$T_{88}(\text{Au}) = 1337.58 \text{ K.} \quad (1)$$

In 1989, an absolute spectroradiometric determination of the temperature of freezing gold was performed at NIST [6] by measuring the spectral radiances of a gold blackbody at wavelengths near 600 nm relative to those of a laser-irradiated integrating sphere which was calibrated with absolute silicon-photodiode detectors and an electrically calibrated radiometer. The result obtained,¹

$$T_{\text{NIST}}(\text{Au}) = (133.33 \pm 0.34) \text{ K,} \quad (2)$$

is 0.25 K smaller than the IPTS-68 value, eq (1) and provided an independent confirmation of measurements by others who had also found a smaller gold-point temperature (see table 3). The NIST gold-point result, eq (2) is identical to the value,

$$T_{90}(\text{Au}) = 1337.33 \text{ K,} \quad (3)$$

which is used as one of the fixed points of the new International Temperature Scale of 1990 (ITS-90) [7].

Effective July 1, 1990, the NIST gold point, eq (2), is used instead of the IPTS-68 gold point, eq (1), for the above-mentioned scale realizations.

¹All uncertainties quoted in this paper are three-standard-deviation (3σ) estimates.

which is used as one of the fixed points of the new International Temperature Scale of 1990 (ITS-90) [7].

Effective July 1, 1990, the NIST gold point, eq (2), is used instead of the PITS-68 gold point, eq (1), for the above-mentioned scale realizations.

Following the adoption of the ITS-90, all other national laboratories having blackbody-based radiometric scales have also implemented similar scale change. The defining text of the ITS-90 which necessitated this change is:

In the radiation temperature region above the freezing point of silver, the ITS-90 can be realized by optical pyrometry and the temperature T_{90} is defined by the equation

$$\frac{L_{\lambda}(T_{90})}{L_{\lambda}[T_{90}(X)]} = \frac{\exp[c_{2,90}/\lambda T_{90}(X)] - 1}{\exp(c_{2,90}/\lambda T_{90}) - 1}, \quad (4)$$

where $T_{90}(X)$ refers to any one of the silver ($T_{90}(\text{Ag}) = 1234.93 \text{ K}$), the gold ($T_{90}(\text{Au}) = 1337.33 \text{ K}$), or the copper ($T_{90}(\text{Cu}) = 1357.77 \text{ K}$) freezing points, where $L_{\lambda}(T_{90})$ and $L_{\lambda}[T_{90}(X)]$ are the spectral concentrations of the radiance of a blackbody at the wavelength (in vacuo) λ at T_{90} and at $T_{90}(X)$, respectively, and where $c_{2,90} = 0.014388 \text{ m}\cdot\text{K}$.

In the International Practical Temperature Scale of 1968, Amended Edition of 1975 (IPTS-68) [5] the temperature T_{68} was defined by a similar equation:

$$\frac{L_{\lambda}(T_{68})}{L_{\lambda}[T_{68}(\text{Au})]} = \frac{\exp[c_{2,68}/\lambda T_{68}(\text{Au})] - 1}{\exp(c_{2,68}/\lambda T_{68}) - 1}, \quad (5)$$

where $T_{68}(\text{Au}) = 1337.58 \text{ K}$ and $c_{2,68} = 0.014388 \text{ m}\cdot\text{K}$.

From the above, the following relationships between the ITS-90 and the IPTS-68 are noted:

(1) The value assigned to the second radiation constant, c_2 , has not been changed,

$$c_{2,90} = c_{2,68} = 0.014388 \text{ m}\cdot\text{K}. \quad (6)$$

(2) The ITS-90 gold-point assignment has been lowered by 0.25 K,

$$T_{90}(\text{Au}) = T_{68}(\text{Au}) - 0.25 \text{ K} = 1337.33 \text{ K}. \quad (7)$$

(3) The ITS-90 above the freezing point of silver (1234.93K) is defined in terms of three fixed points, the freezing points of silver, gold (1337.33K), and copper (1357.77K). This introduces a redundancy of the new scale, whereas all previous scales were unambiguously defined in terms of the gold point alone. However, the text of the

ITS-90 states that "the T_{90} values of the freezing points of silver, gold, and copper are believed to be self consistent to such a degree that the substitution of any one of them in place of one of the other two as the reference temperature $T_{90}(X)$ will not result in significant differences in the measured values of T_{90} ."

The changes in radiometric and photometric calibration values due to the 1968/1990 gold-point change are insignificantly small and have been published by the national laboratories. For example, the calibration values reported by NIST for spectral radiance and irradiance decreased by 0.8% at 250 nm, 0.4% at 500 nm, and 0.1% at 2000 nm. Calibration values for luminous intensity and luminous flux decreased by 0.35% and there was no change in NIST calibrations of color temperature.

2. Effect on CIE Definitions and Recommendations

The activities of the CIE which relate to temperature and could conceivably be affected by the adoption of the ITS-90 fall in the domain of thermal radiation and its applications in radiometry, photometry, and colorimetry. The published CIE output in these areas can be classified into two broad areas: definitions and recommendations. In what follows, it will be shown that none of these definitions and recommendations requires revision because of the introduction of the ITS-90.

2.1 CIE Definitions

2.1.1 General

The output of the CIE contains a large number of definitions of the fundamental concepts, physical quantities, and technical terms used in the areas of light and lighting. These definitions have traditionally been collected in the International Lighting Vocabulary, which is now available in its fourth edition [8] as a joint publication compiled by the CIE in collaboration with the International Electrotechnical Commission (IEC). Insofar as these definitions relate to thermal

radiation, they are given in terms of thermo-dynamic temperatures and are, therefore, independent of the International Temperature Scale and not affected by the adoption of the ITS-90. Specific examples include the definitions of color temperature, correlated color temperature, distribution temperature, radiance temperature, ratio temperature, and similar quantities which are used to assign a single temperature to the spectral power distributions of incandescent or fluorescent lamps, phases of natural daylight, and other sources which are not too different from blackbody sources.

2.1.2 CIE Standard Illuminants

The definitions published by the CIE comprise, as a special class, "standards" in the form of uniquely defined data which form the basis of internationally accepted standard systems. These include "standard illuminants" for applications of colorimetry requiring the use of spectral power distributions which are representative of typical lighting conditions [9,10]. These illuminants are spectral power distributions which are not necessarily realizable by laboratory sources, and are provided by the CIE in the form of numerical tables. At the present time, there are two CIE standard illuminants.

2.1.2.1 Standard Illuminant A

CIE Standard Illuminant A is representative of domestic tungsten lighting and was originally defined as follows [11].

"It is recommended that the following ... luminous sources be adopted for the general colorimetry of materials:

"Source A. A gas-filled lamp of color temperature 2848 K ...

"For calculations of the spectral energy distribution the constant, c_2 , of Planck is taken equal to 14,350 micron degrees.

"Source A. The spectral energy distribution of this source shall be considered for all colorimetric applications as that of a blackbody at the temperature 2848 K."

This definition was independent of the International Temperature Scale of 1927, which was in effect at the time. After the introduction of the International Temperature Scale of 1948, and once again after the IPTS-68 was introduced, the CIE followed a policy of adopting the assigned values of c_2 but preserving the relative spectral distribution of illuminant A. This was achieved by shifting the color temperature assigned to it, rather than computing new spectral distributions for the original color temperature of 2848 K. Accordingly, the spectral values of CIE standard illuminant A are to be considered independent of the International Temperature Scale and require no adjustments in view of the adoption of the ITS-90. The color temperature assigned to standard illuminant A is a descriptive parameter which depends on the value of c_2 assigned in the International Temperature Scale. Its value is approximately 2856 K on the IPTS-68 and remains the same on the ITS-90 because the value of c_2 has not been changed.

2.1.2.2 Standard Illuminant D_{65}

This standard is representative of average daylight and is given in the form of numerical values that have been derived by Judd, MacAdam, and Wyszecki [12] from experimental data taken by others. The tabulated spectral data of illuminant D_{65} are consistent with the IPTS-68 value of c_2 , and are also not affected by the introduction of the ITS-90 because c_2 was not changed. Likewise, the correlated color temperature of illuminant D_{65} , which is given as 6504 K on the IPTS-68, remains unchanged.

2.1.3 Other CIE Illuminants

The CIE has also defined illuminants which do not have the status of primary CIE standards but are useful for special purposes [10]. Illuminants B (now obsolete) and C represent direct sunlight and average daylight with correlated color temperatures of approximately 4874 K and 6800 K, respectively. Three illuminants, D_{50} , D_{55} , and D_{75} , have been defined in addition to standard illuminant D_{65} to represent daylight with approximate correlated color temperatures of 5000, 5500, and 7500 K. Additionally, illuminants F_1 through F_{12} have been defined to represent typical fluorescent lamps. The published relative spectral distributions of these additional illuminants [10] are consistent, within their estimated accuracies, with IPTS-68 temperatures and, therefore, with ITS-90 temperatures as well.

2.2 CIE Recommendations

2.2.1 General

The CIE recommendations [10] concerning the calculation of color temperature and related quantities are dependent on the International Temperature Scale, in that a numerical value of c_2 is usually specified in these recommendations. However, the change from IPTS-68 to ITS-90 temperatures does not affect any of these recommendations because the value of c_2 remains the same in the new scale.

2.2.2 Recommended Sources for Producing CIE Illuminants

The CIE has recommended specific sources [9,10] which can be used for practical realizations of the spectral distributions defined by CIE illuminants.

For example, the CIE recommends that Standard Illuminant A be realized by means of a gas-filled tungsten-filament lamp known as "Source A" which is operated at a correlated color temperature (2856 K on the IPTS-68) equal to the color temperature associated with Standard Illuminant A. This definition requires no change with the introduction of the ITS-90 because the color temperature associated with Standard Illuminant A has not changed.

At present, no source is recommended for realizing CIE Standard Illuminant D_{55} . However, the CIE has recommended a method [13] for assessing sources intended for this purpose. The description of this method does not rely on the International Temperature Scale and, therefore, requires no revision.

Illuminants B and C be realized by combining Source A with specially formulated liquid filters [10]. The specifications of these filters are independent of the International Temperature scale and require no revision.

There are no CIE recommendations for realizing the D and F illuminants mentioned under 2.1.3.

2.3 Effects on Calibrated Sources

Although none of the CIE recommendations for laboratory realizations of CIE Standard Illuminants are affected by the adoption of the ITS-90, it should be noted that sources that have previously been calibrated in accordance with these recommendations do not exactly produce the standard spectral distributions. Because the ITS-90 gold-point assignment is closer to the thermodynamic temperature of this fixed point, a source calibrated on the IPTS-68 and operated at a given radiance temperature T_{88} will have the ITS-90 kelvin temperature [14]

$$T_{90} = T_{88} - 0.25 * (T_{88}/1337.33)^2 \quad (8)$$

and will therefore produce a skewed spectral distribution. For example, a Source-A type lamp having the required correlated color temperature of 2856 K on the IPTS-68 will operate at the ITS-90 temperature 2855 K. Its spectral distribution, normalized to 100.00 at 560 nm, will be 0.07% lower than that of Standard Illuminant A at 460 nm, and 0.05% higher at 660 nm. These differences are small compared to the state-of-the-art calibration uncertainties, so that no adjustments or corrections will be necessary in practice.

References

- [1] Walker, J. H., Saunders, R. D., and Hattenburg, A. T., Spectral Radiance Calibrations, NBS Special Publication SP 250-1, U.S. Government Printing Office, Washington, DC 1987.
- [2] Waters, W. R., Walker, J. H., and Hattenburg, A. T., Radiance Temperature Calibrations, NBS Special Publication SP 250-7, U.S. Government Printing Office, Washington, DC 1987.
- [3] Walker, J. H., Saunders, R. D., Jackson, J. K., and McSparron, D. A., Spectral Irradiance Calibrations, NBS Special Publication SP 250-20, U.S. Government Printing Office, Washington, DC 1987.
- [4] Booker, R. L., and McSparron, D. A., Photometric Calibrations, NBS Special Publication SP250-15, U.S. Government Printing Office, Washington, DC 1987.
- [5] Preston-Thomas, H., Metrologia 12, (1976).
- [6] Mielenz, K. D., Saunders, R. D., and Shumaker, J. B., J. Res. Natl. Inst. Stand. Techn. (U.S.) 95, 49 (1990).
- [7] Preston-Thomas, H., Metrologia 27, 3 (1990)
- [8] CIE Publication 17.4, International Lighting Vocabulary, 1987.
- [9] CIE Publication S001, Colorimetric Illuminants, 1986a.
- [10] CIE Publication 15.2, Colorimetry, Second Edition, 1986b.
- [11] CIE. Recueil des Travaux et Comptes Rendu des Seances, 8th Session, Cambridge, England. pp. 19-22. 1931
- [12] Judd, D. B., MacAdam, D. L., and Wyszecki, G., J. Opt. Soc. Am. 54, 1031 (1964)
- [13] CIE Publication 51, A Method for Assessing the Quality of Daylight Simulators for Colorimetry, 1981
- [14] Mielenz, K. D., Saunders, R. D., and Shumaker, J. B., J. Res. Natl. Inst. Stand. Techn. (U.S.) 95, 49 (1990)