



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0
Scope Revised: 2008-09-24

NVLAP Code: 20/D05
Length & Diameter - Indicators

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
up to 0.100	21	M&TE
0.100 to 0.250	82	M&TE
0.250 to 2	82	M&TE

NVLAP Code: 20/D05
Length - Air Amplifiers

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
0.0003 to 0.003	12	M&TE Dimensionair®

NVLAP Code: 20/D05
Length

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
0.0003 to 0.003	13	M&TE All Mahr Federal Inc. AMR Kits

NVLAP Code: 20/D05
Length

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
< 400 arc seconds	0.40 arc seconds	M&TE Electronic Levels System
0 in to 1 in	58 μin	M&TE 400 B3 & B4 Calibrators

NVLAP Code: 20/D05
Length & Diameter - Outside Micrometers

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
0 to 1	58	M&TE
1 to 2	58	M&TE

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0
Scope Revised: 2008-09-24

2 to 3	58	M&TE
3 to 4	58	M&TE
4 to 5	58	M&TE
5 to 6	58	M&TE

NVLAP Code: 20/D05
Length – Field Service Calibration

Range	Best Uncertainty (\pm) ^{note 1}	Remarks
Universal Length Measuring Machines 0.5 in to 12 in	5.0 μ in (0.127 μ m)	Gage Blocks
Universal Height Measuring Machines 5 mm to 700 mm	80 μ in (2.0 μ m)	Calibrated Step Gage
Universal Calibrators 0.5 in	9.0 μ in (229 μ m)	Gage Blocks
Comparators 0.002 in	3.1 μ in (0.079 μ m)	Gage Blocks

NVLAP Code: 20/D09
Roundness

Range	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
0.124 in to 2 in Dia. with a roundness <100 μ in	1 μ in	
0.124 in to 14.5 in Dia. with a roundness \leq 0.004 in	3.5 μ in (0.089 μ m)	
0.124 in to 14.5 in Dia. with a roundness > 0.004 to 0.40 in	25 μ in (0.64 μ m)	

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0
Scope Revised: 2008-09-24

NVLAP Code: 20/D11
Spherical Diameter; Plug

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
up to 1	6	
1 to 2	7	
2 to 4	10	
4 to 10	(10 + 1L)	

NVLAP Code: 20/D11
Ring Gages

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
0.125 to 5.0	7	Mahr 828 CIM
up to 1	6	
1 to 2	7	
2 to 4	10	
4 to 14	(10 + 1L)	

NVLAP Code: 20/D11
Air Rings

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
< 2	18	M&TE
2 to 4	25	M&TE

NVLAP Code: 20/D11
Air Plugs

Range in inches	Best Uncertainty (\pm) in μin ^{note 1}	Remarks
< 1	12	M&TE
\geq 1 to 2	26	M&TE
> 2 to 3	28	M&TE
> 3 to 4	32	M&TE
> 4 to 5	33	M&TE

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0

Scope Revised: 2008-09-24

NVLAP Code: 20/D12

Surface Texture

<i>Range</i>	<i>Best Uncertainty (±) in μin^{note 1}</i>	<i>Remarks</i>
20 μin R _a to 300 μin R _a	2	

NVLAP Code: 20/D12

Surface Finish / Contour Measuring Machines – Field Service Calibration

<i>Range</i>	<i>Best Uncertainty (±)^{note 1}</i>	<i>Remarks</i>
R _a 100 μin to 150 μin	1.18 μin (0.03 μm)	Surface Finish Standard
W _t <60 μin/in.	3.15 μin (0.08 μm)	Straight Edge
Displacement 180 μin to 240 μin	3.0 μin (0.076 μm)	Step Height Standard
Probe Calibration Steps 1 mm to 70 mm	15.8 μin (0.40 μm)	Gage Blocks
Gage Pin Radius 2 mm to 4 mm	5.12 μin (0.13 μm)	Calibrated Gage Pin
Sphere Radius >4mm to 25 mm	5.12 μin (0.13 μm)	Calibrated Sphere (2 ball master)

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0
Scope Revised: 2008-09-24

NVLAP Code: 20/D15
Precision Geometry / Form Measuring Machines

<i>Range</i>	<i>Best Uncertainty (±) ^{note 1}</i>	<i>Remarks</i>
Concentricity		
≤ 14.5 in Dia. and ≤ 13.75 in. Hgt. with a concentricity of ≤ 0.004 in	12 μin (0.3 μm)	
≤ 14.5 in Dia. and ≤ 13.75 in. Hgt with a concentricity of > 0.004 in to 0.040 in to 0.0040 in	27 μin (0.69 μm)	
Cylindricity		
≤ 1.0 in Hgt. and ≤ 14.5 in Dia. with a cylindricity of ≤ 0.0001 in	5 μin (0.13 μm)	
≤ 4.0 in Hgt. and ≤ 14.5 in Dia. with a cylindricity of ≤ 0.004 in	15 μin (0.38 μm)	
> 4.0 in to 13.75 in Hgt. and ≤ 14.5 in Dia. with a cylindricity of ≤ 0.004 in	25 μin (0.64 μm)	
≤ 4.0 in Hgt. and ≤ 14.5 in Dia. with a cylindricity of > 0.004 in to 0.040 in to 0.040 in	29 μin (0.74 μm)	
≥ 4.0 in to 13.75 in Hgt. and ≤ 14.5 in Dia. with a cylindricity of > 0.004 in to 0.040 in A	35 μin (0.89 μm)	
Flatness		
≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a flatness of ≤ 0.004 in	3 μin (0.08 μm)	
≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a flatness of > 0.004 in to 0.040 in	25 μin (0.64 μm)	
Parallelism		
≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a parallelism of ≤ 0.004 in	4 μin (0.10 μm)	
≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a parallelism of > 0.004 in to 0.040 in	25 μin (0.64 μm)	

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



National Voluntary Laboratory Accreditation Program



CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0
Scope Revised: 2008-09-24

Perpendicularity

≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a Perpendicularity of ≤ 0.004 in	5 μin (0.13 μm)
≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a perpendicularity of > 0.004 in to 0.040 in	25 μin (0.64 μm)

Runout

≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a runout of ≤ 0.004 in	4 μin (0.1 μm)
≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a runout of > 0.004 in to 0.040 in	25 μin (0.64 μm)

Total Runout

≤ 14.5 in Dia. ≤ 13.75 in Hgt. with a total runout of ≤ 0.004 in	330 μin (8.4 μm)
--	---------------------

NVLAP Code: 20/D15

Geometry / Form Measuring Machines – Field Service Calibration

Range	Best Uncertainty (±) ^{note 1}	Remarks
Radial Departure <50 μin	1.85 μin (0.047 μm)	Precision Sphere
Axial Deviation <50 μin	1.2 μin (0.03 μm)	Optical Flat
Coning Error <10 μin / in	1.17 μin (0.03 μm)	Precision Sphere

Probe Calibration

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology



**National Voluntary
Laboratory Accreditation Program**



CALIBRATION LABORATORIES

NVLAP LAB CODE 200605-0

Scope Revised: 2008-09-24

<0.040 in	26.4 μin (0.67 μm)	Gage Blocks
Straightness <2 μm / 100 mm	5.9 μin (0.15 μm)	Straight Edge
Z Axis Parallelism <10 μm / m	87.4 μin (2.22 μm)	Cylindrical Square
X Axis Perpendicular <10 μm / m	281 μin (7.74 μm)	Straight Edge

1. Represents an expanded uncertainty using a coverage factor, k = 2, at an approximate level of confidence of 95 %.
2. L in inches
3. L in mm

2008-04-01 through 2009-03-31

Effective dates

Sally S. Bruce

For the National Institute of Standards and Technology