

EP-8 Evoked Potential Device SEPTEMBER 20, 2000

DEC - 8 2000

Section E - 510(k) Summary

This summary of 510(k) safety and effectiveness information is being submitted in accordance with the requirements of the SMDA 1990 and 21 CFR § 807.92.

Name:

Cameron Mahon

Director of R & D

Address:

XLTEK

2568 Bristol Circle Oakville, Ontario Canada, L6H 5S1

Telephone:

(905) 829-5300

Fax:

(905) 829-5304

E-mail:

research@XLTEK.com

Common Names:

Evoked Potential Stimulator

Classification Name:

21 CFR § 882.1870, Evoked Potential

Stimulator, 84 GWF, Class II

21 CFR § 882.189, Evoked Response Stimulator,

84 GWE, Class II

21 CFR § 882.19, Auditory Evoked Response

Stimulator, 84 GWJ, Class II

21 CFR § 882.1400, Electroencepahlograph, 84

GWQ, Class II

21 CFR § 890.1374, Electromyograph,

Diagnostic, 84IKN, Class II

Predicate Devices:

XLTEK EP-16 [510(k) #K992313]

Description:

The XLTEK EP-8 is an 8 channel evoked

potential stimulator which has embedded



EP-8 Evoked Potential Device SEPTEMBER 20, 2000

controls, is computer based, and can be used in conjunction with video and networking.

Substantial Equivalence:

The XLTEK EP-8 is substantially equivalent in terms of safety and effectiveness to the XLTEK EP-16 [510(k) #K992313].

Indications for Use:

The EP-8 is intended for use in the operating room and critical care areas for neurological monitoring and assessment. The instrument uses EEG, evoked potentials, and EMG techniques to provide the health professionals with information to help assess a patient's neurological status during surgery or long term monitoring in the ICU. The XLTEK EP-8 is an 8 channel evoked potential stimulator which has embedded controls, is computer based, and can be used in conjunction with video and networking.



Food and Drug Administration 9200 Corporate Boulevard Rockville MD 20850

DEC - 8 2000

Ms. Debbie Davy
'Research and Development Administration
Excel Tech, Ltd.
2568 Bristol Circle
Oakville, Ontario
Canada L6H 5S1

Re: K002969

Trade Name: XLTEK Evoked Potential Headbox, Model EP-8

Regulatory Class: II

Product Code: GWF, GWE, GWJ, GWQ, IKN

Dated: September 21, 2000 Received: September 22, 2000

Dear Ms. Davy:

We have reviewed your Section 510(k) notification of intent to market the device referenced above and we have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (Premarket Approval), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 895. A substantially equivalent determination assumes compliance with the current Good Manufacturing Practice requirement, as set forth in the Quality System Regulation (QS) for Medical Devices: General regulation (21 CFR Part 820) and that, through periodic (QS) inspections, the Food and Drug Administration (FDA) will verify such assumptions. Failure to comply with the GMP regulation may result in regulatory action. In addition, FDA may publish further announcements concerning your device in the Federal Register. Please note: this response to your premarket notification submission does not affect any obligation you might have under sections 531 through 542 of the Act for devices under the Electronic Product Radiation Control provisions, or other Federal laws or regulations.

This letter will allow you to begin marketing your device as described in your 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801 and additionally 809.10 for in vitro diagnostic devices), please contact the Office of Compliance at (301) 594-4595. Additionally, for questions on the promotion and advertising of your device, please contact the Office of Compliance at (301) 594-4639. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR 807.97). Other general information on your responsibilities under the Act may be obtained from the Division of Small Manufacturers Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its internet address "http://www.fda.gov/cdrh/dsmamain.html".

Sincerely yours,

May Mulherse

Celia M. Witten, Ph.D., M.D.

Director

Division of General, Restorative and Neurological Devices Office of Device Evaluation Center for Devices and Radiological Health

Enclosure



EP-8 Evoked Potential Device SEPTEMBER 20, 2000

Section D – Statement of Indications for Use

Page 1 of 1

510(k) Number (if known):	K002969
Device Name :	XLTEK EP-8 Evoked Potential Headbox
Indications for Use :	The EP-8 is intended for use in the operating room and critical care areas for neurological monitoring and assessment. The instrument uses EEG, evoked potentials, and EMG techniques to provide the health professionals with information to help assess a patient's neurological status during surgery or long term monitoring in the ICU. The XLTEK EP-8 is an 8 channel evoked potential stimulator which has embedded controls, is computer based, and can be used in conjunction with video and networking.
(PLEASE DO NOT WRITE PAGE IF NEEDED)	E BELOW THIS LINE - CONTINUE ON ANOTHER
	f CDRH, Office of Device Evaluation (ODE)
	Mul Muleuses Division Sign-Off
ì	Division Sign-Off) Division of General Rustorative Devices K 002969 S10(k) Number
Prescription Use V	OR Over-The Counter Use

(Per 21 CFR 801.109)