



National Voluntary Laboratory Accreditation Program



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2005

GeoCorp Industrial Controls, Inc.

9010 River Road

Huron, OH 44839

Mr. Niel Fleetwood

Phone: 419-433-1101 Fax: 419-433-1102

E-mail: nfleetwood@geocorpinc.com

URL: <http://www.geocorpinc.com>

CALIBRATION LABORATORIES

NVLAP LAB CODE 200496-0

THERMODYNAMICS

NVLAP Code: 20/T08

Thermocouples

<i>Range in °C</i>	<i>Best Uncertainty (±) in °C^{note 1}</i>	<i>Remarks</i>
0 to 1093	1.50	Type B
> 1093 to 1454	2.24	Type B
0 to 1000	1.50	Type E
0 to 760	1.50	Type J
0 to 1093	1.50	Type K
> 1093 to 1327	2.23	Type K
0 to 1093	1.50	Type N
> 1093 to 1300	2.23	Type N
0 to 1093	1.50	Type R
> 1093 to 1454	2.24	Type R
0 to 1093	1.50	Type S
> 1093 to 1454	2.24	Type S
0 to 400	1.50	Type T

1. Represents an expanded uncertainty using a coverage factor, $k = 2$, at an approximate level of confidence of 95 %.

2008-10-01 through 2009-09-30

Effective dates

For the National Institute of Standards and Technology